

QIAamp[®] DNA Mini QIAcube Kit (240)

For use on the QIAcube[®]

Kit Contents

QIAamp DNA Mini QIAcube Kit (240)	
Catalog no.	51326
QIAamp Mini Spin Columns	240
Rotor Adapters	240
Elution Tubes	240
Buffer AL*	2 x 33 ml
Buffer ATL	50 ml
Buffer AW1* (concentrate)	98 ml
Buffer AW2 [†] (concentrate)	66 ml
Buffer AE	60 ml
Proteinase K	6 ml
Product sheet	1

* Contains chaotropic salt. Not compatible with disinfecting agents containing bleach. † Contains sodium azide as a preservative.

Shipping and Storage

QIAamp Mini Spin Columns and buffers can be stored dry at room temperature (15–25°C) for up to 1 year. Proteinase K solution can be stored at room temperature or at 2–8°C for up to 1 year without affecting performance.



Sample & Assay Technologies

Intended Use

The QIAamp DNA Mini QIAcube Kit is intended for molecular biology applications. This product is not intended for the diagnosis, prevention, or treatment of a disease. For more information about the automated procedure, see the relevant protocol sheet available at www.qiagen.com/MyQIAcube.

Safety Information

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available online in convenient and compact PDF format at www.qiagen.com/Support/MSDS.aspx where you can find, view, and print the SDS for each QIAGEN® kit and kit components.

CAUTION: DO NOT add bleach or acidic solutions directly to the sample-preparation waste.

The sample-preparation waste contains guanidine hydrochloride from Buffers AL and AW1, which can form highly reactive compounds when combined with bleach. If liquid containing these buffers is spilt, clean with suitable laboratory detergent and water. If the spilt liquid contains potentially infectious agents, clean the affected area first with laboratory detergent and water, and then with 1% (v/v) sodium hypochlorite.

24-hour emergency information

Emergency medical information can be obtained 24 hours a day in English, French, and German from: Poison Information Center Mainz, Germany. Tel: +49-6131-19240.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Trademarks: QIAGEN®, QIAamp®, QIAcube® (QIAGEN Group).

1072901 06/2012 © 2012 QIAGEN, all rights reserved.

www.qiagen.com

Canada = 800-572-9613

Ireland = 1800 555 049

Norway = 800-18859

China = 8621-3865-3865

Italy = 800-787980

Singapore = 1800-742-4368

Denmark = 80-885945

Japan = 03-6890-7300

Spain = 91-630-7050

Australia = 1-800-243-800

Finland = 0800-914416

Korea (South) = 080-000-7145

Sweden = 020-790282

Austria = 0800/281010

France = 01-60-920-930

Luxembourg = 8002 2076

Switzerland = 055-254-22-11

Belgium = 0800-79612

Germany = 02103-29-12000

Mexico = 01-800-7742-436

UK = 01293-422-911

Brazil = 0800-557779

Hong Kong = 800 933 965

The Netherlands = 0800 0229592

USA = 800-426-8157



Sample & Assay Technologies