

# QIAamp<sup>®</sup> DNA Blood Mini Accessory Sets

For use with multiple QIAamp DNA Blood Mini Kits on the QIAcube<sup>™</sup>

## Set Contents

QIAamp DNA Blood Mini Accessory Set	A	B
<b>Catalog no.</b>	<b>1043368</b>	<b>1043369</b>
Buffer AL	33 ml	33 ml
Buffer AW1 (concentrate)	19 ml	–
Buffer AW2 (concentrate)	13 ml	–
Buffer AE	60 ml	60 ml
QIAGEN <sup>®</sup> Protease	1 vial	1 vial
Protease Solvent	1.2 ml	1.2 ml
Product Sheet	1	1

## Storage

All buffers can be stored at room temperature (15–25°C) for up to 1 year without showing any reduction in performance. Lyophilized QIAGEN Protease can be stored at room temperature (15–25°C) for up to 12 months without affecting performance. Reconstituted QIAGEN Protease is stable for up to 1 year when stored at 2–8°C. Keeping the QIAGEN Protease stock solution at room temperature for prolonged periods of time should be avoided.

## Product Use Limitations

The QIAamp DNA Blood Mini Accessory Set A contains additional buffers and reagents typically for use with **10 x QIAamp DNA Blood Mini Kits (50)**, catalog number 51104, on the QIAcube. The QIAamp DNA Blood Mini Accessory Set B is typically for use with **2 x QIAamp DNA Blood Mini Kits (250)**, catalog number 51106, on the QIAcube. Depending on run parameters, the set may be sufficient for use with more or, occasionally, fewer complete kits. For setting up the QIAcube for DNA purification using the QIAamp DNA Blood Mini Kit, follow the instructions in the Protocol Sheet, which can be downloaded from [www.qiagen.com/myQIAcube](http://www.qiagen.com/myQIAcube).

QIAamp DNA Blood Mini Accessory Sets are intended for molecular biology applications. These products are not intended for the diagnosis, prevention, or treatment of disease.



## Safety Information

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate material safety data sheets (MSDSs). These are available online in convenient and compact PDF format at [www.qiagen.com/ts/msds.asp](http://www.qiagen.com/ts/msds.asp) where you can find, view, and print the MSDS for each QIAGEN kit and kit component.

**CAUTION: DO NOT add bleach or acidic solutions directly to the sample-preparation waste.**

Buffers AL and AW1 contain guanidine hydrochloride, which can form highly reactive compounds when combined with bleach. If liquid containing these buffers is spilt, clean with suitable laboratory detergent and water. If the spilt liquid contains potentially infectious agents, clean the affected area first with laboratory detergent and water, and then with 1% (v/v) sodium hypochlorite.

The following risk and safety phrases apply to components of QIAamp DNA Blood Mini Accessory Sets:

### Buffer AL and Buffer AW1

Contains guanidine hydrochloride: harmful, irritant. Risk and safety phrases:\* R22-36/38, S13-26-36-46

### QIAGEN Protease

Contains subtilisin: sensitizer, irritant. Risk and safety phrases:\* R37/38-41-42, S22-24-26-36/37/39-46

### 24-hour emergency information

Emergency medical information in English, French, and German can be obtained 24 hours a day from:

Poison Information Center Mainz, Germany

Tel: +49-6131-19240

\* R22: Harmful if swallowed; R36/38: Irritating to eyes and skin; R37/38: Irritating to respiratory system and skin; R41: Risk of serious damage to eyes; R42: May cause sensitization by inhalation; S13: Keep away from food, drink and animal feedingstuffs; S22: Do not breathe dust; S24: Avoid contact with skin; S26: In case of contact with eyes, rinse immediately with plenty of water and seek medical advice; S36: Wear suitable protective clothing; S36/37/39: Wear suitable protective clothing, gloves and eye/face protection; S46: If swallowed, seek medical advice immediately and show the container or label.

Trademarks: QIAGEN®, QIAamp®, QIAcube™ (QIAGEN Group).

1063050 04/2010 © 2007 QIAGEN, all rights reserved.

[www.qiagen.com](http://www.qiagen.com)

Australia = 1800-243-800

Finland = 0800-914416

Korea (South) = 080-000-7146

Sweden = 020-790282

Austria = 0800-28-10-10

France = 01-60-920-920

Luxembourg = 8002-2076

Switzerland = 055-254-22-11

Belgium = 0800-79612

Germany = 02103-29-12000

Mexico = 01-800-7742-639

UK = 01293-422-911

Brazil = 55-11-5079-4000

Hong Kong = 800-933-965

The Netherlands = 0800-0229592

USA = 800-426-8157

Canada = 800-572-9613

Ireland = 1800-555-049

Norway = 800-18859

China = 800-988-0326

Italy = 02-33430411

Singapore = 1800-742-4362

Denmark = 80-885945

Japan = 03-5547-0811

Spain = +34-91-630-7050



Sample & Assay Technologies