

QIASymphony SP Protocol Sheet

DNA_Buffy_Coat_400_V6_DSP protocol

General information

For in vitro diagnostic use.

This protocol is for purification of total genomic and mitochondrial DNA from fresh or frozen buffy coat using the QIASymphony® SP and the QIASymphony DSP DNA Midi Kit.

Kit	QIASymphony DSP DNA Midi Kit (cat. no. 937255)
Sample material	Buffy coat (EDTA, citrate, or heparin anti-coagulated)
Protocol name	DNA_BC_400_V6_DSP
Default Assay Control Set	ACS_BC_400_V6_DSP
Editable	Elution volume: 200 μ l, 400 μ l
Required software version	Version 4.0

September 2012



Sample & Assay Technologies

“Sample” drawer

Sample type	Buffy coat (EDTA, citrate, or heparin anti-coagulated)
Sample volume	Depends on type of sample tube used; for more information see www.qiagen.com/goto/dsphandbooks .
Primary sample tubes	n/a
Secondary sample tubes	For more information see www.qiagen.com/goto/dsphandbooks .
Inserts	Depends on type of sample tube used; for more information see www.qiagen.com/goto/dsphandbooks .

n/a = not applicable.

“Reagents and Consumables” drawer

Position A1 and/or A2	Reagent cartridge
Position B1	n/a
Tip rack holder 1–17	Disposable filter-tips, 200 µl or 1500 µl
Unit box holder 1–4	Unit boxes containing sample prep cartridges or 8-Rod Covers

n/a = not applicable.

“Waste” drawer

Unit box holder 1–4	Empty unit boxes
Waste bag holder	Waste bag
Liquid waste bottle holder	Empty liquid waste bottle

“Eluate” drawer

Elution rack (we recommend using slot 1, cooling position)	For more information see www.qiagen.com/goto/dsphandbooks .
---	---

Required plasticware

	One batch, 24 samples*	Two batches, 48 samples*	Three batches, 72 samples*	Four batches, 96 samples*
Disposable filter-tips, 200 μ l ^{††}	4	4	4	8
Disposable filter-tips, 1500 μ l ^{††}	110	212	314	424
Sample prep cartridges [§]	18	36	54	72
8-Rod Covers [¶]	3	6	9	12

* Use of less than 24 samples per batch decreases the number of disposable filter-tips required per run.

† There are 32 filter-tips/tip rack.

† Number of required filter-tips includes filter-tips for 1 inventory scan per reagent cartridge.

§ There are 28 sample prep cartridges/unit box.

¶ There are twelve 8-Rod Covers/unit box.

Note: Numbers of filter-tips given may differ from the numbers displayed in the touchscreen depending on settings. We recommend loading the maximum possible number of tips.

Elution volume

The elution volume is selected in the touchscreen. Depending on the sample type and DNA content, the final eluate volume may vary by up to 15 μ l less than the selected volume. Due to the fact that the eluate volume may vary, we recommend checking the actual eluate volume when using an automated assay setup system that does not verify the eluate volume prior to transfer. Elution in lower volumes increases the final DNA concentration, but slightly reduces the yield. We recommend using an elution volume appropriate for the intended downstream application.

Preparation of sample material

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, consult the appropriate safety data sheets (SDSs), available from the product supplier.

Important point before starting

- QIAAsymphony magnetic particles may copurify RNA if it is present in the sample. In order to minimize RNA content in the sample, add RNase A to the sample before starting the procedure. The final RNase A concentration should be 2 mg/ml.

Buffy coat

Buffy coat is a leukocyte-enriched fraction of whole blood. The efficiency of leukocyte enrichment depends on the procedure used to prepare buffy coat and on the accuracy with which the buffy coat layer is extracted. Prepare buffy coat by centrifuging whole blood samples containing a standard anticoagulant (EDTA, citrate, or heparin) at 900–1100 x g for 10 minutes at room temperature (15–25°C). After centrifugation, 3 different fractions are distinguishable: the upper clear layer is plasma; the intermediate layer is buffy coat, containing concentrated leukocytes; and the bottom layer contains concentrated erythrocytes. Approximately 1 ml leukocyte-containing fraction should be harvested from 10 ml centrifuged whole blood, which, on average, gives a 5–6x enrichment. For example, 10 ml whole blood with a white blood cell count of 6×10^6 cells/ml results in 1 ml buffy coat. Assuming a 5x enrichment of white blood cells, this results in 3×10^7 cells/ml. Therefore, in a protocol that uses 400 μ l buffy coat, 1.2×10^7 cells will be used.

To avoid overloading the DNA purification procedure, do not prepare buffy coat samples of >10x enrichment. If buffy coat samples are of >10x enrichment, dilute the samples to 10x enrichment or less with PBS or use less starting material in the DNA purification procedure.

Buffy coat samples may be used immediately or stored at –20°C or –70°C for purification of DNA at a later date. Frozen samples should be thawed quickly in a 37°C water bath with mild agitation to ensure thorough mixing and then equilibrated to room temperature (15–25°C) before beginning the procedure. To ensure reliable sample transfer, avoid generating foam in sample tubes. Try to avoid blood clots in the samples and, if necessary, transfer the sample without clots to a fresh tube.

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Trademarks: QIAGEN®, QIAAsymphony® (QIAGEN Group). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.
© 2012 QIAGEN, all rights reserved.

www.qiagen.com
Australia ■ 1-800-243-800
Austria ■ 0800/281010
Belgium ■ 0800-79612
Canada ■ 800-572-9613
China ■ 021-51345678
Denmark ■ 80-885945
Finland ■ 0800-914416

France ■ 01-60-920-930
Germany ■ 02103-29-12000
Hong Kong ■ 800 933 965
Ireland ■ 1800 555 049
Italy ■ 800 787980
Japan ■ 03-5547-0811
Korea (South) ■ 1544 7145
Luxembourg ■ 8002 2076

The Netherlands ■ 0800 0229592
Norway ■ 800-18859
Singapore ■ 65-67775366
Spain ■ 91-630-7050
Sweden ■ 020-790282
Switzerland ■ 055-254-22-11
UK ■ 01293-422-911
USA ■ 800-426-8157

