

March 2021

artus[®] CMV RG PCR Kit

Instructions for Use (Handbook)



24 (catalog no. 4503263)



96 (catalog no. 4503265)

Version 1

Quantitative in vitro diagnostic

For use with Rotor-Gene[®] Q MDx Instruments

IVD

CE 0197

REF

4503263, 4503265



QIAGEN GmbH

QIAGEN Strasse 1, 40724 Hilden, GERMANY

R6 **MAT**

1123965

Contents

Intended Use	5
Description and Principle	5
Pathogen Information	6
Principle of the Procedure	6
Materials Provided	7
Kit Contents	7
Materials Required but Not Provided	8
Reagents	8
Consumables	8
Equipment	8
Warnings and Precautions	9
Safety information	9
Precautions	9
Reagent Storage and Handling	10
Specimen Handling and Storage	10
Specimen Collection	10
Sample Storage	11
Sample Transport	11
Procedure	12
DNA Isolation	12
Internal Control	13
Protocol: PCR and Data Analysis	14

Interpretation of Results.....	22
Quantitation	22
Results.....	23
Quality Control.....	26
Limitations.....	26
Performance Characteristics	27
Analytical Sensitivity	27
Linear Range.....	29
Specificity.....	30
Precision	32
Interfering Substances	34
Robustness.....	36
Reproducibility	36
Diagnostic Evaluation	38
References	40
Troubleshooting Guide	41
Symbols.....	43
Ordering Information	44
Document Revision History	47

Intended Use

The *artus* CMV RG PCR Kit is an in vitro nucleic acid amplification test for the quantitation of cytomegalovirus (CMV) DNA in human plasma. This diagnostic test kit utilizes the polymerase chain reaction (PCR) and is configured for use with Rotor-Gene Q Instruments.

The *artus* CMV RG PCR Kit is intended for use in conjunction with clinical presentation and other laboratory markers for the management of CMV infection in patients at risk for CMV disease.

The results from the *artus* CMV RG PCR Kit must be interpreted within the context of all relevant clinical and laboratory findings.

The *artus* CMV RG PCR Kit is not intended for use as a screening test for the presence of CMV in blood or blood products or as a diagnostic test to confirm the presence of CMV infection.

Description and Principle

The *artus* CMV RG PCR Kit constitutes a ready-to-use system for the detection of CMV DNA using polymerase chain reaction (PCR) on Rotor-Gene Q MDx Instruments. The CMV RG Master contains reagents and enzymes for the specific amplification of a 105 bp region of the Major Immediate Early Gene (*MIE*) within the CMV genome (the assay is able to detect CMV genotypes gB1 – gB4), and for the direct detection of this specific amplicon in fluorescence channel Cycling Green of the Rotor-Gene Q MDx.

In addition, the *artus* CMV RG PCR Kit contains a second heterologous amplification system to identify possible PCR inhibition. This is detected as an internal control (IC) in fluorescence channel Cycling Yellow of the Rotor-Gene Q MDx. External positive controls (CMV QS 1–4) are supplied, which allow the determination of the amount of viral DNA. For further information, see “Quantitation”, page 22.

Pathogen Information

The human cytomegalovirus (CMV) is found in infected persons in blood, tissues, and nearly all secretory fluids. Transmission can be oral, sexual, via blood transfusion or organ transplantation, intrauterine, or perinatal (1-4). CMV viral load testing is an important aid for assessing risk of disease, diagnosing disease, and monitoring response to therapy (5).


Infection with CMV frequently leads to an asymptomatic infection followed by a lifelong persistence of the virus in the body. If symptoms occur in teenagers or in adults, they resemble those of mononucleosis with fever, weak hepatitis, and general indisposition (6). Severe courses of CMV infection have been observed especially in those infected intrauterine and in immunodeficient patients (4,7).

Principle of the Procedure

Pathogen detection by polymerase chain reaction (PCR) is based on the amplification of specific regions of the pathogen genome. In real-time PCR, the amplified product is detected via fluorescent dyes. These are usually linked to oligonucleotide probes that bind specifically to the amplified product. Monitoring the fluorescence intensities during the PCR run (i.e., in real time) allows the detection and quantitation of the accumulating product without having to reopen the reaction tubes after the PCR run (8).

Materials Provided

Kit Contents

<i>artus</i> CMV RG PCR Kit		(24)	(96)
Catalog no.		4503263	4503265
Number of reactions		24	96
Blue	CMV RG Master (Taq 0.1 U/μl)		2 x 12 reactions
Yellow	CMV Mg-Sol*	Mg-Sol	600 μl
Red	CMV QS 1† (1 x 104 copies/μl)	QS	200 μl
Red	CMV QS 2† (1 x 103 copies/μl)	QS	200 μl
Red	CMV QS 3† (1 x 102 copies/μl)	QS	200 μl
Red	CMV QS 4† (1 x 101 copies/μl)	QS	200 μl
Green	CMV RG IC‡	IC	1000 μl
White	Water (PCR grade)		1000 μl
	Instructions for use		1

* Magnesium solution

† Quantitation standard

‡ Internal control

Materials Required but Not Provided

Reagents

- DNA isolation kit (see “DNA Isolation”, page 12)

Consumables

- Sterile pipette tips with filters
- Strip Tubes and Caps, 0.1 ml, for use with 72-well rotor (cat. no. 981103 or 981106)
- **Alternatively:** PCR Tubes, 0.2 ml, for use with 36-well rotor (cat. no. 981005 or 981008)

Equipment

- Pipettes (adjustable)*
- Vortex mixer*
- Benchtop centrifuge* with rotor for 2 ml reaction tubes
- Rotor-Gene Q MDx Instruments* with fluorescence channels for Cycling Green and Cycling Yellow
- Rotor-Gene Q Software version 2.3.5 or higher
- Cooling block (Loading Block 72 x 0.1 ml Tubes, cat. no. 9018901, or Loading Block 96 x 0.2 ml Tubes, cat. no. 9018905)

* Prior to use, ensure that instruments have been checked and calibrated according to the manufacturer's recommendations.

Warnings and Precautions

Safety information

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, please consult the appropriate safety data sheets (SDSs). These are available online in convenient and compact PDF format at www.qiagen.com/safety where you can find, view, and print the SDS for each QIAGEN kit and kit component.

Discard sample and assay waste according to your local safety regulations.

Precautions

The user should always pay attention to the following:

- Use sterile pipette tips with filters.
- Store and extract positive materials (specimens, positive controls, and amplicons) separately from all other reagents, and add them to the reaction mix in a spatially separated facility.
- Thaw all components thoroughly at room temperature (15–25°C) before starting an assay.
- When thawed, mix the components (by pipetting repeatedly up and down or by pulse vortexing) and centrifuge briefly.
- Work quickly and keep components on ice or in the cooling block (72/96 well loading block).

Reagent Storage and Handling

The components of the *artus* CMV RG PCR Kit should be stored at -30°C to -15°C and are stable until the expiration date stated on the label. Repeated thawing and freezing ($>2x$) should be avoided, as this may reduce assay sensitivity. If the reagents are to be used intermittently, they should be frozen in aliquots. Storage at $2-8^{\circ}\text{C}$ should not exceed a period of 5 hours.

Specimen Handling and Storage

Note: All samples must be treated as potentially infectious material.

Note: Analytical studies carried out to verify performance of this kit refer to EDTA plasma as the most suitable sample materials for CMV detection. Therefore, we recommend the use of this material with the *artus* CMV RG PCR Kit.

The validation of the *artus* CMV RG PCR Kit has been performed using human EDTA plasma samples. Other sample materials are not validated. Please use only the recommended nucleic acid isolation kit (see "DNA Isolation", page 12) for sample preparation.

When using certain sample materials, particular instructions regarding collection, transport, and storage must be strictly observed.

Specimen Collection

Each blood withdrawal causes an injury of blood vessels (arteries, veins, or capillaries). Only innocuous and sterile material should be used. For blood withdrawal, appropriate disposables should be available. For vein punctures, capillary needles that are too fine should not be employed. Venous blood withdrawal should be carried out on the appropriate parts of the elbow bend, the forearm, or the back of the hand. Blood has to be withdrawn with standard specimen collection tubes (red cap, Sarstedt® or equivalent tube of another manufacturer). A volume of 5–10 ml blood should be withdrawn into an EDTA tube. Tubes should be mixed overhead directly after sample collection (8x, do not agitate).

Note: Heparinized samples must not be used.

Sample Storage

Whole blood should be separated into plasma and cellular components by centrifugation for 20 minutes at 800–1600 $\times g$ within 6 hours (9,10). The isolated plasma is transferred into sterile polypropylene tubes. The sensitivity of the assay can be reduced if samples are routinely frozen or stored for a longer period of time.

Sample Transport

Sample material should be transported in a shatterproof transport container as a matter of principle. Thus, a potential danger of infection due to a leakage of sample can be avoided. The samples should be transported following the local and national instructions for the transport of pathogen material.*

The samples should be shipped within 6 hours. It is not recommended to store the samples where they have been collected. It is possible to ship the samples by mail, following the legal instructions for the transport of pathogen material. We recommend sample transport with a courier. The blood samples should be shipped cooled (2–8°C), and the separated plasma deep frozen (–30 to –15°C).

* International Air Transport Association (IATA). Dangerous Goods Regulations.

Procedure

DNA Isolation

The kits from QIAGEN shown in Table 1 are validated for viral DNA purification from the indicated human sample types for use with the *artus* CMV RG PCR Kit. Carry out the viral DNA purification according to the instructions in the respective kit handbooks.

Table 1. Purification kits validated for use with the *artus* CMV RG PCR Kit

Sample material	Sample size	Nucleic acid isolation kit	Catalog number	Carrier RNA
EDTA plasma	500 µl	QIAamp® DSP Virus Kit	60704	Included
EDTA plasma	400 µl	EZ1® DSP Virus Kit (48)	62724	Included

Note: The use of carrier RNA is critical for the extraction efficiency and, consequently, for DNA/RNA yield. To increase the stability of the carrier RNA provided with the QIAamp DSP Virus Kit, we recommend proceeding according to the information about the reconstitution and storage of the carrier RNA given in the “Preparing reagents and buffers” section of the *QIAamp DSP Virus Kit Handbook*.

Note: The internal control of the *artus* CMV RG PCR Kit can be used directly in the isolation procedure. Make sure to include one negative plasma sample in the isolation procedure. The corresponding signal of the internal control is the basis for the evaluation of the isolation (see “Internal Control” section below).

Internal Control

An internal control (CMV RG IC) is supplied with this kit. This allows the user to both control the DNA isolation procedure and to check for possible PCR inhibition. For this application, add the internal control to the isolation at a ratio of 0.1 μl per 1 μl elution volume. For example, using the QIAamp DSP Virus Kit, the DNA is eluted in 60 μl Elution Buffer (AVE). Hence, 6 μl of the internal control should be added initially. The quantity of internal control used depends only on the elution volume.

Note: The internal control and carrier RNA (see “DNA Isolation”, page 12) should be added only to the mixture of lysis buffer and sample material or directly to the lysis buffer.

The internal control must not be added to the sample material directly. If added to the lysis buffer, please note that the mixture of internal control and lysis buffer–carrier RNA must be freshly prepared and used immediately (storage of the mixture at room temperature or in the fridge for only a few hours may lead to internal control failure and a reduced extraction efficiency).

Note: Do not directly add the internal control and the carrier RNA to the sample material.

To consider a purification successful, the C_T value of the internal control of a negative plasma sample that has been processed during purification (QIAamp DSP Virus Kit) has to reach $C_T = 27 \pm 3$ (threshold: 0.03) using Rotor-Gene Q Instruments (see page 25 for more information). The stated spreading is based on the variance of the instrument and the purification. A higher deviation points to a purification problem. In this case, the purification has to be checked and, if necessary, validated a second time. If you have any further questions or if you encounter problems, contact QIAGEN Technical Services.

Optionally, the internal control can be used exclusively to check for possible PCR inhibition. For this application, add the internal control directly to the CMV RG Master and CMV Mg-Sol, as described in step 2b of the protocol (page 15).

Protocol: PCR and Data Analysis

Important points before starting

- Take time to familiarize yourself with the Rotor-Gene Q Instrument before starting the protocol. See the respective instrument user manual for more information.
- Make sure that at least one quantitation standard as well as one negative control (Water, PCR grade) are included per PCR run. To generate a standard curve, use all 4 quantitation standards supplied (CMV QS 1–4) for each PCR run.

Things to do before starting

- Make sure that the cooling block (accessory of the Rotor-Gene Q Instrument) is precooled to 2–8°C.
- Before each use, all reagents need to be thawed completely, mixed (by repeated up and down pipetting or by quick vortexing), and centrifuged briefly.

Procedure

1. Place the desired number of PCR tubes into the adapters of the cooling block.
2. If you are using the internal control to monitor the DNA isolation procedure and to check for possible PCR inhibition, follow step 2a. If you are using the internal control exclusively to check for PCR inhibition, follow step 2b.

Note: It is highly recommended to add the internal control to CMV RG Master and CMV Mg-Sol used for the quantitation standards. For the quantitation standards, add the internal control directly to the CMV RG Master and CMV Mg-Sol, as described in step 2b of the protocol, and use this master mix for each quantitation standard (CMV QS 1–4).

- 2a. The internal control has already been added to the isolation (see *Internal Control*, page 13). In this case, prepare a master mix according to Table 2 (next page). The reaction mix typically contains all of the components needed for PCR, except the sample.

Table 2. Preparation of master mix (internal control used to monitor DNA isolation and check for PCR inhibition)

Number of samples	1	12
CMV RG Master	25 µl	300 µl
CMV Mg-Sol	5 µl	60 µl
CMV RG IC	0 µl	0 µl
Total volume	30 µl	360 µl

2b. The internal control must be added directly to the mixture of CMV RG Master and CMV Mg-Sol. In this case, prepare a master mix according to Table 3.

The reaction mix typically contains all of the components needed for PCR, except the sample.

Table 3. Preparation of master mix (internal control used exclusively to check for PCR inhibition)

Number of samples	1	12
CMV RG Master	25 µl	300 µl
CMV Mg-Sol	5 µl	60 µl
CMV RG IC	2 µl	24 µl
Total volume	32 µl*	384 µl*

* The volume increase caused by adding the internal control is neglected when preparing the PCR assay. The sensitivity of the detection system is not impaired.

3. Pipet 30 µl of the master mix into each PCR tube, then add 20 µl of the eluted sample DNA (see Table 4). Correspondingly, 20 µl of at least one of the quantitation standards (CMV QS 1–4) must be used as a positive control and 20 µl of water (Water, PCR grade) as a negative control.

Table 4. Preparation of PCR assay

Number of samples	1	12
Master mix	30 µl	30 µl each
Sample	20 µl	20 µl each
Total volume	50 µl	50 µl each
Number of samples	1	12

4. Close the PCR tubes. Make sure that the locking ring (accessory of the Rotor-Gene Instrument) is placed on top of the rotor to prevent accidental opening of the tubes during the run.
5. For the detection of CMV DNA, create a temperature profile according to the following steps.

Setting the general assay parameters	Figure 1, Figure 2, and Figure 3
Initial activation of the hot-start enzyme	Figure 4
Amplification of the DNA (touchdown PCR)	Figure 5
Adjusting the fluorescence channel sensitivity	Figure 6
Starting the run	Figure 7

All specifications refer to the Rotor-Gene Q Software version 2.3.5 or higher. Please find further information on programming Rotor-Gene Instruments in the respective instrument user manual. In the illustrations, these settings are framed in bold black. Illustrations are included for Rotor-Gene Q Instruments.

6. Open the **New Run Wizard** dialog box (Figure 1, next page). Tick the **Locking Ring Attached** box and click **Next**.




Figure 1. The “New Run Wizard” dialog box.

7. Select 50 for the PCR reaction volume and click **Next** (Figure 2).




Figure 2. Setting the general assay parameters.

- Click the **Edit Profile** button in the next **New Run Wizard** dialog box (Figure 3), and program the temperature profile as shown in Figure 3 to Figure 5).




Figure 3. Editing the profile.




Figure 4. Initial activation of the hot-start enzyme.




Figure 5. Amplification of the DNA. Make sure to activate the touchdown function for 10 cycles in the annealing step.

9. The detection range of the fluorescence channels has to be determined according to the fluorescence intensities in the PCR tubes. Click **Gain Optimisation** in the **New Run Wizard** dialog box (see Figure 3, previous page) to open the **Auto-Gain Optimisation Setup** dialog box. Set the calibration temperature to 65°C to match the annealing temperature of the amplification program (Figure 6, next page).




Figure 6. Adjusting the fluorescence channel sensitivity.

- The gain values determined by the channel calibration are saved automatically and are listed in the last menu window of the programming procedure (Figure 7, next page). Click **Start Run**.




Figure 7. Starting the run.

Interpretation of Results

Quantitation

The enclosed quantitation standards (CMV QS 1–4) are treated as previously purified samples, and the same 20 µl volume is used directly in the PCR (no need for further extraction). To generate a standard curve on Rotor-Gene Q Instruments, all 4 quantitation standards should be used and defined in the **Edit Samples** dialog box as standards with the specified concentrations (see the respective instrument user manual).

Note: To ensure accurate quantitation, it is highly recommended to add the internal control to CMV RG Master and CMV Mg-Sol used for the quantitation standards. For this application, add the internal control directly to the CMV RG Master and CMV Mg-Sol, as described in step 2b of the protocol (page 15), and use this master mix for each quantitation standard (CMV QS 1–4).

Note: The quantitation standards are defined as copies/µl. The following equation has to be applied to convert the values determined using the standard curve into copies/ml of sample material:

$$\text{Result} \left(\frac{\text{copies}}{\text{ml}} \right) = \frac{\text{Result (copies/}\mu\text{l)} \times \text{Elution Volume (}\mu\text{l)}}{\text{Sample Volume (ml)}}$$

As a matter of principle, the initial sample volume should be entered in the equation above. This has to be considered when the sample volume has been changed prior to the nucleic acid extraction (e.g., reducing the volume by centrifugation or increasing the volume by adding to the volume required for the isolation).

Note: The quantitation standards have been calibrated against the 1st International Standard for Human Cytomegalovirus (NIBSC code: 09/162) as determined by the World Health Organization (WHO).

To convert copies/ml to IU/ml in consideration of the QIAamp DSP Virus Kit:

$$\text{WHO (IU/ml)} = 2.933 \times \text{artus CMV (copies/ml)}$$

Note: For the QIAamp workflow, quantified samples must be within the linear range of the QS 1×10^1 to 1×10^4 copies/ μ l. Quantitation cannot be assured outside of this range.

To convert copies/ml to IU/ml in consideration of the EZ1 DSP Virus Kit on the EZ1 Advanced XL instrument:

$$\text{WHO (IU/ml)} = 0.794 \times \text{artus CMV (copies/ml)}$$

Note: For the EZ1 workflow, quantified samples must be within the linear range $3.16E+02$ to $1.00E+08$ copies/ml. Quantitation cannot be assured outside of this range.

Results

Examples of positive and negative PCR reactions are given in Figure 8 and Figure 9, (next page).




Figure 8. Detection of the quantitation standards (CMV QS 1-4) in fluorescence channel Cycling Green. NTC: No template control (negative control).




Figure 9. Detection of the internal control (IC) in fluorescence channel Cycling Yellow with simultaneous amplification of the quantitation standards (CMV QS 1-4). NTC: No template control (negative control).

A signal is detected in fluorescence channel Cycling Green.

The result of the analysis is positive: the sample contains CMV DNA.

In this case, the detection of a signal in the Cycling Yellow channel is dispensable, since high initial concentrations of CMV DNA (positive signal in the Cycling Green channel) can lead to a reduced or absent fluorescence signal of the internal control in the Cycling Yellow channel (competition).

In fluorescence channel Cycling Green no signal is detected. At the same time, a signal from the internal control appears in the Cycling Yellow channel.

No CMV DNA is detectable in the sample. It can be considered negative.

In the case of a negative CMV PCR, the detected signal of the internal control rules out the possibility of PCR inhibition.

No signal is detected in the Cycling Green or in the Cycling Yellow channels.

No result can be concluded.

Information regarding error sources and their solution can be found in the "Troubleshooting Guide", page 41.

Quality Control

In accordance with QIAGEN's ISO-certified Quality Management System, each lot of *artus* CMV RG PCR Kit is tested against predetermined specifications to ensure consistent product quality.

Limitations

All reagents are for in vitro diagnostic use only.

The product is to be used by personnel specially instructed and trained with in vitro diagnostic procedures.

Strict compliance with the respective instrument user manual is required for optimal PCR results.

Attention should be paid to expiration dates printed on the box and labels of all components. Do not use expired components.

Although rare, mutations within the highly conserved regions of the viral genome covered by the kit's primers and/or probe may result in underquantitation or failure to detect the presence of the virus in these cases. Validity and performance of the assay design are revised at regular intervals.

Performance Characteristics

Analytical Sensitivity

The analytical detection limit as well as the analytical detection limit in consideration of the purification (sensitivity limits) were assessed for the *artus* CMV RG PCR Kit. The analytical detection limit in consideration of the purification is determined using CMV-positive clinical specimens in combination with a particular extraction method. In contrast, the analytical detection limit is determined independently from the selected extraction method, using CMV DNA of known concentration.

To determine the analytical sensitivity of the *artus* CMV RG PCR Kit, a dilution series of CMV genomic DNA was set up from 10 to nominal 0.00316 copies/ μ l and analyzed on Rotor-Gene Instruments in combination with the *artus* CMV RG PCR Kit. Testing was carried out on 3 different days on 8 replicates. The results were determined by a probit analysis. A graphical illustration of the probit analysis on the Rotor-Gene 6000 is shown in Figure 10 (next page). The analytical detection limit of the *artus* CMV RG PCR Kit in combination with the Rotor-Gene Q MDx/Q/6000 and the Rotor-Gene 3000 is 0.36 copies/ μ l ($p = 0.05$) and 0.24 copies/ μ l ($p = 0.05$), respectively. This means that there is a 95% probability that 0.36 copies/ μ l or 0.24 copies/ μ l will be detected.




Figure 10. Probit analysis: CMV (Rotor-Gene 6000). Analytical sensitivity of the *artus* CMV RG PCR Kit on the Rotor-Gene 6000.

The analytical sensitivity in consideration of the purification (QIAamp DSP Virus Kit) of the *artus* CMV RG PCR Kit on Rotor-Gene Instruments was determined using a dilution series of CMV virus material from 1000 to nominal 0.316 CMV copies/ml spiked in clinical plasma specimens. These were subjected to DNA extraction using the QIAamp DSP Virus Kit (extraction volume: 0.5 ml, elution volume: 60 μl). Each of the 8 dilutions was analyzed with the *artus* CMV RG PCR Kit on 3 different days on 8 replicates. The results were determined by a probit analysis. A graphical illustration of the probit analysis is shown in Figure 11 (next page). The analytical detection limit in consideration of the purification of the *artus* CMV RG PCR Kit in combination with the Rotor-Gene 3000 is 57.1 copies/ml ($p = 0.05$). This means that there is a 95% probability that 57.1 copies/ml will be detected.




Figure 11. Probit analysis: CMV (Rotor-Gene 3000). Analytical sensitivity in consideration of the purification (QIAamp DSP Virus Kit, QIAGEN) of the *artus* CMV RG PCR Kit on the Rotor-Gene 3000.

The analytical sensitivity in consideration of the purification with EZ1 DSP Virus Kit (extraction volume: 0.4 ml, elution volume: 60 µl) using the EZ1 Advanced XL instrument of the *artus* CMV RG PCR Kit on the Rotor-Gene 6000 is 68.75 copies/ml ($p = 0.05$). This means that there is a 95% probability that 68.75 copies/ml will be detected.

Linear Range

The linear range in consideration of the purification with EZ1 DSP Virus Kit (extraction volume: 0.4ml, elution volume: 60 µl) using the EZ1 Advanced XL instrument was determined by testing 4 to 6 replicates of CMV virus material in a dilution series from $3.16E+01$ to $1.00E+08$ copies/ml.

A graphical illustration of the probit analysis is shown in Figure 12 (next page).




Figure 12. Polynomial regression of dataset of *artus* CMV RG PCR Kit in consideration of the purification (EZ1 DSP Virus Kit) on the EZ1 Advanced XL instrument. Linear, quadratic, and cubic regression models are included.

The linear range of the *artus* CMV RG PCR Kit in consideration of the purification with EZ1 DSP Virus Kit (extraction volume: 0.4ml, elution volume: 60µl) using the EZ1 Advanced XL instrument is 3.16E+02 to 1.00E+08 copies/ml.

Note: The linear range of the *artus* CMV RG PCR Kit, in consideration of the purification with QIAamp DSP Virus Kit (extraction volume: 0.4 ml, elution volume: 60 µl), is 1.00E+01 to 1.00E+04 copies/µl.

Specificity

The specificity of the *artus* CMV RG PCR Kit is first and foremost ensured by the selection of the primers and probes, as well as the selection of stringent reaction conditions. The primers and probes were checked for possible homologies to all sequences published in gene banks by sequence comparison analysis. The detectability of all relevant strains has thus been ensured.

Moreover, the specificity was validated with 100 different CMV negative plasma samples. 99 of these samples did not generate any signals with the CMV specific primers and probes, which are included in the CMV RG Master.

Note: 1 sample that generated a signal in the CMV-specific primers and probes that also tested CMV positive in the *artus* CMV LC and TM RG PCR Kits are likely positive. The final specificity based on testing 100 individual donor samples was verified as 99.00% (99/100).

A potential cross-reactivity of the *artus* CMV RG PCR Kit was tested using the control group listed in Table 5. None of the tested pathogens were reactive. No cross-reactivities appeared with mixed infections.

Table 5. Testing the specificity of the kit with potentially cross-reactive pathogens

Control group	CMV (Cycling Green or Cycling A.FAM)	Internal control (Cycling Yellow or Cycling A.JOE)
Human herpesvirus 1 (Herpes simplex virus 1)	–	+
Human herpesvirus 2 (Herpes simplex virus 2)	–	+
Human herpesvirus 3 (Varicella-zoster virus)	–	+
Human herpesvirus 4 (Epstein-Barr virus)	–	+
Human herpesvirus 6A	–	+
Human herpesvirus 6B	–	+
Human herpesvirus 7	–	+
Human herpesvirus 8 (Kaposi's sarcoma-associated herpesvirus)	–	+
Hepatitis A virus	–	+
Hepatitis B virus	–	+
Hepatitis C virus	–	+

(continued on next page)

Table 5 (continued from previous page)

Control group	CMV (Cycling Green or Cycling A.FAM)	Internal control (Cycling Yellow or Cycling A.JOE)
Human immunodeficiency virus 1	–	+
Human T cell leukemia virus 1	–	+
Human T cell leukemia virus 2	–	+
West Nile virus	–	+
Enterovirus	–	+
Parvovirus B19	–	+

Precision

The precision data of the *artus* CMV RG PCR Kit have been collected by means of Rotor-Gene Instruments and allow the determination of the total variance of the assay. The total variance consists of the intra-assay variability (variability of multiple results of samples of the same concentration within one experiment), the inter-assay variability (variability of multiple results of the assay generated on different instruments of the same type by different operators within one laboratory), and the inter-batch variability (variability of multiple results of the assay using various batches). The data obtained were used to determine the standard deviation, the variance, and the coefficient of variation for the pathogen specific and the internal control PCR.

Precision data of the *artus* CMV RG PCR were collected using the quantitation standard of the lowest concentration (QS 4; 10 copies/ μ l). Testing was performed with 8 replicates. The precision data were calculated on basis of the C_T values of the amplification curves (C_T : threshold cycle, see Table 6, next page). In addition, precision data for quantitative results in copies/ μ l were determined using the corresponding C_T values (see Table 7, next page). Based on these results, the overall statistical spread of any given sample with the mentioned concentration is 1.21% (C_T) or 14.38% (concentration), and 1.93% (C_T) for the detection of the internal control. These values are based on the totality of all single values of the determined variabilities.

Table 6. Precision data on basis of the C_T values

	Standard deviation	Variance	Coefficient of variation (%)
Intra-assay variability: CMV QS 4	0.17	0.03	0.57
Intra-assay variability: Internal control	0.31	0.10	1.16
Inter-assay variability: CMV QS 4	0.38	0.14	1.27
Inter-assay variability: Internal control	0.47	0.22	1.77
Inter-batch variability: CMV QS 4	0.33	0.11	1.10
Inter-batch variability: Internal control	0.53	0.28	2.02
Total variance: CMV QS 4	0.36	0.13	1.21
Total variance: Internal control	0.51	0.26	1.93

Table 7. Precision data on basis of the quantitative results (in copies/ μ l)

	Standard deviation	Variance	Coefficient of variation (%)
Intra-assay variability: CMV QS 4	1.34	1.80	13.30
Inter-assay variability: CMV QS 4	1.54	2.38	15.25
Inter-batch variability: CMV QS 4	1.46	2.12	14.41
Total variance: CMV QS 4	1.45	2.11	14.38

Interfering Substances

CMV DNA was spiked into negative plasma in different commercially available blood collection systems with different anticoagulants. The calculated concentration (copies/ml), C_T mean, standard deviation, variance, and CV% are reported in Table 8. The standard deviation and coefficient of variation are within the scope of 5% and thus within the tolerance range. No significant impact on the PCR due to the various substances was identified.

Table 8. Commercial blood collection systems and anticoagulants data

Substance	Concentration (copies/ml)	C_T mean	C_T Standard deviation	C_T Variance	C_T CV (%)
Potassium EDTA, Becton Dickinson®	399.60	31.06	0.11	0.01	0.36
Potassium EDTA, Sarstedt	350.10	31.26	0.30	0.09	0.97
Potassium EDTA, Greiner Bio-One®	285.00	31.58	0.50	0.25	1.58
Potassium EDTA, Springe (reference)	310.40	31.40	0.16	0.03	0.52
Potassium EDTA, Sarstedt (reference)	487.20	30.80	0.14	0.02	0.47
Potassium EDTA (pregnancy)	423.30	33.2	0.26	0.07	0.79

Endogenous substances (Table 9, next page) were spiked into CMV-positive EDTA plasma samples at 3 x LOD and 10 x LOD. All samples were successfully detected, and no interference was observed for samples containing elevated levels of endogenous inhibitors (bilirubin, hemoglobin, triglyceride and albumin).

Table 9. Endogenous substances tested

Interfering substances	Concentration of interfering substances
Bilirubin	30 mg/dl
Hemoglobin	2 g/dl
Triglyceride	1 g/dl
Albumin	6 g/dl

Common drugs used in transplant settings were tested at 3x the acute peak concentration following a drug therapeutic treatment, as recommended in the CLSI® Guideline EP07-A2 (11) (refer to Table 10). Each of these substances were spiked into both CMV-negative and CMV-positive samples that were tested in 4 replicates.

All tested exogenous substances showed no significant influence on the performance of the *artus* CMV RG PCR kit.

Table 10. List of Drugs tested as exogenous substances

Interfering substances	Test Concentration
Antibiotics	
Sulfamethoxazole	200 mg/l
Trimethoprim	5.2 mg/l
Claforan® (Cefotaxime)	1 g/l
Tazobac® (Piperacillin+Tazobactam)	Piperacillin: 1 g/l Tazobactam: 125 mg/l
Ticarcillin	1 g/l
Augmentin® (Amoxicillin + Clavulanic acid)	Amoxicillin: 125 mg/l Clavulanic acid: 25 mg/l
Vancomycin	125 mg/l
Antifungal	
Fluconazole	1 mg/l
Immunosuppressive drugs	
Rapamycin	100 mg/l
Mycophenolate sodium	80 mg/l

Robustness

The verification of the robustness allows the determination of the total failure rate of the *artus* CMV RG PCR Kit. 100 CMV negative samples of plasma were spiked with CMV at a final concentration of 170 copies/ml (approximately threefold concentration of the analytical sensitivity limit). After extraction using the QIAamp DSP Virus Kit, these samples were analyzed with the *artus* CMV RG PCR Kit. For all CMV samples, the failure rate was 0%. In addition, the robustness of the internal control was assessed by purification and analysis of 100 CMV-negative plasma samples. Thus, the robustness of the *artus* CMV RG PCR Kit is $\geq 99\%$.

Reproducibility

Reproducibility data permit a regular performance assessment of the *artus* CMV RG PCR Kit, as well as an efficiency comparison with other products. These data are obtained by the participation in established proficiency programs.

In addition to participation in established proficiency programs, a 10-member CMV panel (Table 11) was tested across 3 external laboratories using the EZ1 DSP Virus Kit on the EZ1 Advanced XL instrument to purify nucleic acid and the *artus* RG PCR kit to test the DNA eluate.

Table 11. Summary of CMV Panel Members

Panel Number (Panel Member type)	Panel Member	Dilution Effect
1001 (1)	Negative	Negative Pool 1
1002 (1)	Negative	Negative Pool 2
1003 (2)	High Negative	50% Positive
1004 (2)	High Negative	50% Positive
1005 (3)	Low Positive	200 copies/ml
1006 (3)	Low Positive	200 copies/ml
1007 (4)	Moderate Positive	2,000 copies/ml
1008 (4)	Moderate Positive	2,000 copies/ml
1009 (5)	High Positive	200,000 copies/ml
1010 (5)	High Positive	200,000 copies/ml

The 10-member panel was tested in duplicate by 2 different operators each day for 6 days at each site with 3 reagent kit lots. Therefore, 20 samples multiplied by 2 operators for 6 days in 3 sites equals 720 data points.

The total reproducibility of the *artus* CMV RGQ MDx test was found to be $\leq 12\%$ CV for samples with a concentration between 200 copies/ml and 200,000 copies/ml (Table 12)

Table 12. Overall Summary (each Panel Member type) – observed averages

panel_member_type	No of Obs	Mean	Median	Standard Deviation	Percent CV	Minimum
1	144	0.02	0.00	0.158	849.84	0.00
2	144	0.68	0.83	0.630	92.19	-0.10
3	144	1.91	1.95	0.226	11.83	0.98
4	144	2.96	2.96	0.168	5.68	2.16
5	144	5.03	5.03	0.091	1.80	4.75

The overall summary of the percentage variance and standard deviation for the log₁₀ IU/ml values for each of the 5 panels across lot, site, operator, day, between run, and within run is presented in Table 13 (next page).

Table 13. Overall summary of variance and standard deviation

Sample	1	2	3	4	5	
Sample type	negative	high negative	low positive	moderate positive	high positive	
Observed mean log₁₀ IU/ml	0.02	0.68	1.91	2.96	5.03	
No. of tests	144	144	144	144	144	
Measure	%variance S.D.					
Lot	0	3.10	0	0	3.00	
	0	0.113	0	0	0.016	
Site	0	0	0	0.90	0	
	0	0	0	0.016	0	
Operator	4.3	4.6	0	18.8	15.4	
	0.033	0.136	0	0.074	0.037	
Variance Component	Day	0	0	8.60	6.00	48.10
		0	0	0.067	0.042	0.065
	Between run	0	0	4.40	10.90	7.90
		0	0	0.048	0.057	0.026
	Within run	95.7	92.3	87	63.40	25.60
		0.155	0.611	0.212	0.136	0.048
Total	100	100	100	100	100	
	0.158	0.635	0.227	0.171	0.094	

Diagnostic Evaluation

The *artus* CMV RG PCR Kit was evaluated in a study comparing the *artus* CMV RG PCR Kit to the COBAS® AMPLICOR® CMV MONITOR® Test. 156 retrospective and prospective clinical EDTA plasma samples were analyzed. All specimen samples had previously been analyzed positive or negative using the COBAS AMPLICOR CMV MONITOR for routine diagnostics.

CMV DNA for testing the *artus* CMV RG PCR Kit was isolated using the QIAamp DSP Virus Kit, with the internal control of the *artus* CMV RG PCR Kit added to the isolation, and analysis was carried out on the Rotor-Gene 3000. The specimens for the COBAS AMPLICOR CMV MONITOR Test were processed and analyzed according to the instructions of the manufacturer provided in the package insert.

All 11 samples that tested positive with the COBAS AMPLICOR CMV MONITOR Test also tested positive with the *artus* CMV RG PCR Kit. 123 of 145 samples that tested negative with the COBAS AMPLICOR CMV MONITOR Test also tested negative with the *artus* CMV RG PCR Kit. 22 discordant results were obtained (Table 14).

Table 14. Results of the comparative validation study

		COBAS AMPLICOR CMV MONITOR Test		Total
		+	-	
<i>artus</i> CMV RG PCR Kit	+	11	22	33
	-	0	123	123

If the results of the COBAS AMPLICOR CMV MONITOR Test are taken as reference, the diagnostic sensitivity of all samples of the *artus* CMV RG PCR Kit is 100%, and the diagnostic specificity is 84.8%.

Further testing of the 22 discordant samples confirmed the results of the *artus* PCR Kits. Therefore it can be assumed that the discrepancy is based on the higher sensitivity of the *artus* CMV RG PCR Kit.

References

1. Plosa E.J., Esbenshade J.C., Fuller M.P., and Weitkamp J.H. (2012). Cytomegalovirus Infection. *Pediatr. Rev.* **33**, 156-163.
2. Furui Y., Satake M., Hoshi Y., Uchida S., Suzuki K., and Tadokoro K. (2013). Cytomegalovirus (CMV) seroprevalance in Japanese blood donors and high detection frequency of CMV DNA in elderly donors. *Transfusion.* **53**, 2190-2197.
3. Atabani, S.F., et al. (2012). Cytomegalovirus replication kinetics in solid organ transplant recipients managed by preemptive therapy. *Am. J. Transplant.* **12**, 2457-2464.
4. Enders G., Daiminger A., Bäder U., Exler S., and Enders M. (2011). Intrauterine transmission and clinical outcome of 248 pregnancies with primary cytomegalovirus infection in relation to gestational age. *J. Clin. Virol.* **52**, 244-246.
5. Kotton, C.N., et al. (2018). The Third International Consensus Guidelines on the Management of Cytomegalovirus in Solid-organ Transplantation. *Transplantation.* **102**, 900-931.
6. Lancini D, Faddy H.M., Flower R., and Hogan C. (2014). Cytomegalovirus disease in immunocompetent adults. *Med. J. Aust.* **201**, 578-580.
7. Eddleston M, Peacock S, Juniper M, and Warrell DA. (1997). Severe cytomegalovirus infection in immunocompetent patients. *Clin. Infect. Dis.* **24**, 52-56.
8. Mackay, I.M. (2004). Real-time PCR in the microbiology laboratory. *Clin. Microbiol. Infect.* **10**, 190-212.
9. Nesbitt S.E., Cook L., Jerome K.R. (2004). Cytomegalovirus quantitation by real-time PCR is unaffected by delayed separation of plasma from whole blood. **42**, 1296-1297.
10. Abdul-Ali D., Kraft C.S., Ingersoll J., Frempong M., Caliendo A.M. (2011). Cytomegalovirus DNA stability in EDTA Anti-Coagulated Whole Blood and Plasma Samples. *J. Clin. Virol.* **52**, 222-224
11. Clinical and Laboratory Standards Institute (CLSI). (2005). *Interference Testing in Clinical Chemistry: Approved Guideline – Second Edition*. CLSI Document EP07-A2. Wayne, PA: Clinical and Laboratory Standards Institute (formerly NCCLS).

Troubleshooting Guide

This troubleshooting guide may be helpful in solving any problems that may arise. For more information, see also the Frequently Asked Questions page at our Technical Support Center: www.qiagen.com/FAQ/FAQList.aspx.

Comments and suggestions

No signal with positive controls (CMV QS 1–4) in fluorescence channel Cycling Green

- | | |
|---|---|
| a) The selected fluorescence channel for PCR data analysis does not comply with the protocol | For data analysis, select the fluorescence channel Cycling Green for the analytical CMV PCR and the fluorescence channel Cycling Yellow for the internal control PCR. |
| b) Incorrect programming of the temperature profile of the Rotor-Gene Instrument | Compare the temperature profile with the protocol. See "Protocol: PCR and Data Analysis", page 14. |
| c) Incorrect configuration of the PCR | Check your work steps by means of the pipetting scheme, and repeat the PCR, if necessary. See "Protocol: PCR and Data Analysis", page 14. |
| d) The storage conditions for one or more kit components did not comply with the instructions given in "Reagent Storage and Handling", page 10) | Check the storage conditions and the expiration date (see the kit label) of the reagents and use a new kit, if necessary. |
| e) The <i>artus</i> CMV RG PCR Kit has expired | Check the storage conditions and the expiration date (see the kit label) of the reagents and use a new kit, if necessary. |

Weak or no signal of the internal control of a negative plasma sample subjected to purification using the QIAamp DSP Virus Kit ($C_T = 27 \pm 3$; threshold, 0.03) in fluorescence channel Cycling Yellow and simultaneous absence of a signal in channel Cycling Green

- | | |
|---|--|
| a) The PCR conditions do not comply with the protocol | Check the PCR conditions (see above) and repeat the PCR with corrected settings, if necessary. |
| b) The PCR was inhibited | Make sure that you use the recommended isolation method and closely follow the manufacturer's instructions. |
| c) DNA was lost during extraction | If the internal control was added to the extraction, an absent signal of the internal control can indicate the loss of DNA during the extraction. Make sure that you use the recommended isolation method (see "DNA Isolation", page 12) and closely follow the manufacturer's instructions. |

Comments and suggestions

- | | |
|---|---|
| d) The storage conditions for one or more kit components did not comply with the instructions given in "Reagent Storage and Handling", page 10) | Check the storage conditions and the expiration date (see the kit label) of the reagents and use a new kit, if necessary. |
| e) The <i>artus</i> CMV RG PCR Kit has expired | Check the storage conditions and the expiration date (see the kit label) of the reagents and use a new kit, if necessary. |

Signals with the negative controls in fluorescence channel Cycling Green of the analytical PCR

- | | |
|---|---|
| a) Contamination occurred during preparation of the PCR | Repeat the PCR with new reagents in replicates.
If possible, close the PCR tubes directly after addition of the sample to be tested.
Make sure to pipet the positive controls last.
Make sure that work space and instruments are decontaminated at regular intervals. |
| b) Contamination occurred during extraction | Repeat the extraction and PCR of the sample to be tested using new reagents.
Make sure that work space and instruments are decontaminated at regular intervals. |

Symbols



Contains reagents sufficient for <N> tests



Use by



In vitro diagnostic medical device



Catalog number



Lot number



Material number



Components



Contains



Number



Global Trade Item Number



Temperature limitation



Manufacturer



Consult instructions for use

Ordering Information

Product	Contents	Cat. no.
<i>artus</i> CMV RG PCR Kit (24)	For 24 reactions: Master, Magnesium Solution, 4 Quantitation Standards, Internal Control, Water (PCR grade)	4503263
<i>artus</i> CMV RG PCR Kit (96)	For 96 reactions: Master, Magnesium Solution, 4 Quantitation Standards, Internal Control, Water (PCR grade)	4503265
EZ1 DSP Virus Kit – for automated, simultaneous purification of viral DNA and RNA from 1–14 serum, plasma, or CSF samples		
EZ1 DSP Virus Kit (48)	For 48 viral nucleic acid preps: Prefilled Reagent Cartridges, Disposable Tip Holders, Disposable Filter-Tips, Sample Tubes, Elution Tubes, Buffers, Carrier RNA	62724
QIAamp DSP Virus Kit – for purification of viral nucleic acids from human plasma for in vitro diagnostic purposes		
QIAamp DSP Virus Kit	For 50 preps: QIAamp MinElute® Spin Columns, Buffers, Reagents, Tubes, Column Extenders, and VacConnectors	60704
Rotor-Gene Q MDx and accessories		
Rotor-Gene Q MDx 5plex Platform	Real-time PCR cycler with 5 channels (green, yellow, orange, red, crimson), laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training not included	9002022
Rotor-Gene Q MDx 5plex System	Real-time PCR cycler with 5 channels (green, yellow, orange, red, crimson), laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training	9002023
Rotor-Gene Q MDx 5plex HRM Platform	Real-time PCR cycler and High Resolution Melt analyzer with 5 channels (green, yellow, orange, red, crimson) plus HRM channel, laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training not included	9002032

Product	Contents	Cat. no.
Rotor-Gene Q MDx 5plex HRM System	Real-time PCR cycler and High Resolution Melt analyzer with 5 channels (green, yellow, orange, red, crimson) plus HRM channel, laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training	9002033
Rotor-Gene Q MDx 6plex Platform	Real-time PCR instrument with 6 channels (blue, green, yellow, orange, red, crimson), including laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training not included	9002042
Rotor-Gene Q MDx 6plex System	Real-time PCR instrument with 6 channels (blue, green, yellow, orange, red, crimson), including laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training	9002043
Rotor-Gene Q MDx 2plex Platform	Real-time PCR cycler with 2 channels (green, yellow), laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training not included	9002002
Rotor-Gene Q MDx 2plex System	Real-time PCR cycler with 2 channels (green, yellow), laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training	9002003
Rotor-Gene Q MDx 2plex HRM Platform	Real-time PCR cycler and High Resolution Melt analyzer with 2 channels (green, yellow) plus HRM channel, laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training not included	9002012
Rotor-Gene Q MDx 2plex HRM System	Real-time PCR cycler and High Resolution Melt analyzer with 2 channels (green, yellow) plus HRM channel, laptop computer, software, accessories: includes 1-year warranty on parts and labor, installation and training	9002013
Loading Block 72 x 0.1 ml Tubes	Aluminum block for manual reaction setup with a single-channel pipet in 72 x 0.1 ml tubes	9018901

Product	Contents	Cat. no.
Loading Block 96 x 0.2 ml Tubes	Aluminum block for manual reaction set-up in a standard 8 x 12 array using 96 x 0.2 ml tubes	9018905
Strip Tubes and Caps, 0.1 ml (250)	250 strips of 4 tubes and caps for 1000 reactions	981103
Strip Tubes and Caps, 0.1 ml (2500)	10 x 250 strips of 4 tubes and caps for 10,000 reactions	981106
PCR Tubes, 0.2 ml (1000)	1000 thin-walled tubes for 1000 reactions	981005
PCR Tubes, 0.2 ml (10000)	10 x 1000 thin-walled tubes for 10,000 reactions	981008

For up-to-date licensing information and product-specific disclaimers, see the respective QIAGEN kit handbook or user manual. QIAGEN kit handbooks and user manuals are available at www.qiagen.com or can be requested from QIAGEN Technical Services or your local distributor.

Document Revision History

Revision	Changes
R6, March 2021	<p>Addition of Linear Range, Interfering Substances, and Reproducibility sections.</p> <p>Updates to the Intended Use and Quantitation sections.</p> <p>Removal of references to QIAGEN instruments that are no longer supported.</p>

Limited License Agreement for *artus* CMV RG PCR Kit

Use of this product signifies the agreement of any purchaser or user of the product to the following terms:

1. The product may be used solely in accordance with the protocols provided with the product and this handbook and for use with components contained in the kit only. QIAGEN grants no license under any of its intellectual property to use or incorporate the enclosed components of this kit with any components not included within this kit except as described in the protocols provided with the product, this handbook, and additional protocols available at www.qiagen.com. Some of these additional protocols have been provided by QIAGEN users for QIAGEN users. These protocols have not been thoroughly tested or optimized by QIAGEN. QIAGEN neither guarantees them nor warrants that they do not infringe the rights of third-parties.
2. Other than expressly stated licenses, QIAGEN makes no warranty that this kit and/or its use[s] do not infringe the rights of third-parties.
3. This kit and its components are licensed for one-time use and may not be reused, refurbished, or resold.
4. QIAGEN specifically disclaims any other licenses, expressed or implied other than those expressly stated.
5. The purchaser and user of the kit agree not to take or permit anyone else to take any steps that could lead to or facilitate any acts prohibited above. QIAGEN may enforce the prohibitions of this Limited License Agreement in any Court, and shall recover all its investigative and Court costs, including attorney fees, in any action to enforce this Limited License Agreement or any of its intellectual property rights relating to the kit and/or its components.

For updated license terms, see www.qiagen.com.

The purchase of this product allows the purchaser to use it for the performance of diagnostic services for human in vitro diagnostics. No general patent or other license of any kind other than this specific right of use from purchase is granted hereby.

Trademarks: QIAGEN[®], Sample to Insight[®], QIAamp[®], *artus*[®], EZ1[®], MinElute[®], Rotor-Gene[®] (QIAGEN Group); CLSI[®], (Clinical Laboratory and Standards, Inc.); Augmentin[®] (Glaxo Group Limited); Tazobac[®] (Pfizer Inc.); AMPLICOR[®], COBAS[®], MONITOR[®] (Roche Group); Claforan (Sanofi-Aventis Group); FAM[™], JOE[™] (Thermo Fisher Scientific).

HB-0046-008 1123965 R6 03/2021© 2021 QIAGEN, all rights reserved.

Ordering www.qiagen.com/shop | Technical Support support.qiagen.com | Website www.qiagen.com